

## airmoBTX/VOC C6C12-Calib

For instruments with previous electronic boards until February 2020

### Preventive maintenance kits



		Designation	Item number	Qty
PM Kit One year	CS_CO_C6C12-KIT1	Coalescent filter	CS_FI_00208-0000	1
		O-Ring 1.5 x 0.75 mm Nitril	AR_TU_09002-0000	20
		O-Ring Gasket 4 x 1	AR_TU_09007-0000	2
		O-Ring Gasket 6 x 1	AR_TU_09104-0000	1
		Ignitor Assy FID New	AR_EL_05019-0000	1
		Rotor 6 ports 1/8" HT	CS_PN_00005-0106	1
		Membran and valves Kit for external and internal airmoPUMP	EP_SA_00004-0001	1
PM Kit Two year	CS_CO_C6C12-KIT2	Permeation tube Benzene - around 15 ng/min at 40°C (airmotec certified at ± 10%)	CS_CH_01100-2014	1
		O-ring OR 22.5 x 1.5	CS_TU_09000-0000	1
PM Kit 3 years	CS_CO_C6C12-KIT3	Trap for airmoVOC C6-C12 or BTEX	AR_SA_05157-0000	1
		Compl. Nozzle	AR_ME_05018-0000	1
		Pneumatic actuator 6 port	CS_PN_00005-0002	1
		24V Distributor for pneumatic valve - (fittings included)	CS_SE_00004-0024	1
		Hard disk 128Go SSD 2,5 (SATA connection) since 03/2012	IT_CP_00340-0128	1
PM Kit 5 years	CS_CO_C6C12-KIT5	Column Oven Fan 24V DC 119*119 mm	CS_SE_05015-0001	1
		CPU Board, tested, Incl. Memory supply and H8	CS_CT_01000-CPUT	1

For more details, please check our website (pictures, references, prices, spare parts kits...)

<https://support.chromatotec.com/>